



New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,
Milton Keynes, Bucks MK17 0PD
Tel: 01296 714477
Fax: 01296 714483
E-mail: support@newcityscales.co.uk
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 10154

DATE: 20th March 2021

QUALITY ASSURANCE

CERTIFICATE OF TEST

CUSTOMER: Sharp Skips Ltd

YOUR PURCHASE ORDER:

ADDRESS: Ferry Lane
Rainham
Essex
RM13 9HA


TYPE:
Vishay VT300

SERIAL NO
T16310304

CAPACITY:
50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures, and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED: 
For and on behalf of New City Scales.

POSITION: A.P.

DATE: 20th March 2021





New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,
Milton Keynes, Bucks MK17 0PD
Tel: 01296 714477
Fax: 01296 714483
E-mail: support@newcityscales.co.uk
Web Site: www.newcityscales.co.uk

CALIBRATION RECORD

TYPE: VT300

MANUFACTURE: Vishay

SERIAL: T16310304

CAPACITY / INCREMENTS: 50.000kg x 20kg

ALLOWABLE TOLERANCE + - 20kg


TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
2,000kg	2,000kg	_____
4,000kg	4,000kg	_____
6,000kg	6,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
18,000kg	18,000kg	_____
20,000kg	20,000kg	_____
Substituted with dead-weight	19,840kg	_____
29,840kg	29,840kg	_____
39,840kg	39,840kg	_____
43,700kg	43,700kg	_____

Individual load-cell test. No 1 = 16,000kg. 2 = 16,000kg. 3 = 16,000kg. 4 = 16,000kg.

CALIBRATION WEIGHTS USED: Supplied by New City Scales Cert no: TR11149 / TR11150

TESTED BY NEW CITY SCALES LTD

SIGNED 

DATED : 20th March 2021

